

JUN 24 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130174</u> ✓	
Name <u>Gary Howe</u>		ID# <u>7219</u> ✓	Date <u>6/17/10</u> ✓
A	Agency <u>Haines P.D.</u>	Phone # <u>766-2121</u>	
Instrument Location <u>Haines P.D.</u> ✓			
Alco S/N <u>X124631</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>600</u>
B	Alco Test Values		
	<u>.082</u> ✓ 1 st Alco	<u>.082</u> ✓ 2 nd Alco	
Signature <u>Gary Howe</u>			
(OVER)			

BMB
6/24/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of July, 2010.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130174

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

JUNE 17, 2010

OPERATOR'S NAME:

LOWE/GARY/A

OPERATOR'S NUMBER: 7219

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 12345

DEPT/AGENCY: MNSI

CASE/REPORT: 06172010

TEST TYPE: U

ALCO TARGET VALUE: .082

ALCO S/N: X124631

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.40 in
ALCO TARGET .083 14:02
BLANK TEST .000 14:03
INTERNAL STANDARD VERIFIED 14:03
ALCO TV 30.40 in .082 14:03
BLANK TEST .000 14:04
SUBJECT SAMPLE .000 14:04
BLANK TEST .000 14:05
ALCO TV 30.40 in .082 14:05
BLANK TEST .000 14:06

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

JUNE 17, 2010

TIME 14:07

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS
SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.40 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~